



AS COMPRESSOR NEBULIZER

**Content Of the Box:** 

## 2. INSTRUCTION

for easy inhalation.

Thank you for purchasing the Dr.Odin Compressor Nebulizer. It is a compact medical device designed to efficiently deliver physician prescribed medication to the bronchial lung passages. With proper care and use, it will provide you with many years of reliable treatment.

This product is developed for the successful treatment of asthma, allergies and other respiratory dis-orders. It creates a stream of air that travels through clear tube to the nebulizer. When air enters the nebulizer, it will convert the prescribed medication into aerosol mist

Dr.Odin Compressor Nebulizer should be used under the supervision of a licensed physician and/or a respiratory therapist. We encourage you to thoroughly read product other than its intended use should always be

# 3. IMPORTANT SAFEGUARDS Note: Read all instructions

carefully before use. The following basic precautions are needed when using an electrical product:

### Caution: Failure to read and observe all precautions could result in personal injury or equipment damage.

This unit is not suitable for use in anaesthetic breathing systems This unit is not suitable in suspension or high viscosity form. In such cases information should be sought from the drug supplier.

## **Product Cautions:**

8. Disposal:

 To avoid electrical shock: Keep the unit away from water. Do not immerse the power cord to the unit in liquid. Do not use while bathing. Do not reach for a unit that has fallen into water-immediately unplug the unit.

2. Never operate the unit if it has any damaged parts (including power cord), if it has been dropped or submersed in water. Immediately ask to the technical Person for 3. The unit should not be used where flammable gas, oxygen or aerosol spray are being

4. Disconnect the unit from the electrical outlet before cleaning.

- 5. Do not use other attachments unless recommended by manufacturer. **Operation Cautions**
- . Connect this product to an appropriate voltage outlet for your model.
- Do not run this product unattended
   Never operate if this unit has a damaged cord or plug, if it has been dropped or in

any way into water, if it does not work properly. Return it to a service center for repair

4. If any abnormality occurs, discontinue use immediately until the unit has been examined and repaired.

i. Always unplug the product immediately after use.

6. Never block the air openings of the main unit or place it where the air openings may

## Storage Cautions:

. Do not store the unit under direct sunlight. High temperature or humidity. Keep the unit out of reach of small children.

### Cleaning Cautions:

3. Do not immerse the unit in water. It may damage the unit.

4. Disconnect the unit from the electrical outlet before cleaning i. Clean all necessary parts after each use as instructed in this guidebook.

## 4. OPERATING YOUR COMPRESSOR NEBULIZER

Note: Prior to initial operation, the nebulizer should be thoroughly cleaned.

1. Place your compressor Nebulizer on a flat and stable surface. Be sure that you can

easily reach the controls when you are seated. 2. Open the clear cover and remove the inside accessories.

Important: Prior to initial operation, the nebulizer should be fully cleaned.

Referring to the cleaning procedures in the manual.

3. Gently twist the top part of the nebulizer counter-clockwise to disassemble

4. Fill the bottom section of the nebulizer with the medication prescribed by your physician.

Be sure that the cone is put inside the bottom section. And the marking "6cc" marked on the nebulizer is the maximum level mark. And the medication can not be

filled beyond the maximum level. 5. Gently twist the top part clockwise to reassemble the nebulizer. Be sure the two sections fit well.

6. Attach one end of air tube to the base of the nebulizer. 7. Attach the other end of air tube to the air tube connector

located on the front of the compressor. 8. Attach mouth piece or mask per your choice to the top section of

nebulizer.

9. Plug the power cord into an appropriate electrical outlet. Make sure at this stage, the power switch is at 'O' status.

10. Press power switch to begin your prescribed treatment.

Important: The compressor motor has a thermal protector which will shut off the unit before the unit is overheated. When the thermal protector shuts the unit off, please:

- a. Switch off the unit b. Unplug the unit from the electrical outlet.
- C. Wait about 30 minutes for the motor to cool down before another

Make sure the air openings are not obstructed.

11. When treatment is finished, shut off the unit and unplug it from the electrical outlet.

# 5. CLEANING

It is recommended that the nebulizer, mouthpiece and mask are properly cleaned with hot water after each use and cleaned with a mild detergent after the last treatment of the day. If your physician or respiratory therapist specifies a different cleaning procedure, follow their instructions. Rinsing (after each treatment)

- Disconnect the air tube, nebulizer, mouthpiece and mask.
- 2. Gently twist the nebulizer to open it. 3. Rinse the nebulizer, mouthpiece and mask with water.
- 4.Dry them with clean soft towel or let it air dry.
- 5. Reassemble the nebulizer when completely dry and put these parts in a dry sealed container.

Please follow the following steps to disinfect your nebulizer unless otherwise specified by your physician. It is suggested that the unit is disinfected after the last treatment of the day.

1. Use one part white vinegar with three parts distilled water. Make sure mixed solution is enough to submerge the nebulizer, mouthpiece and mask. 2. Complete rinsing steps 1-3. 3. Wash nebulizer, mouthpiece and mask in warm water and a mild detergent.

Then wash them in hot tap water. 4. Submerge these parts in the vinegar and water solution for thirty minutes. 5.

Complete rinsing steps 3-5. Cleaning the compressor:

1. Wipe daily with a damp cloth.

2. Do not use any powdered cleaners or soap pads, which may damage the

1. Do not use cotton or any other materials. Do not wash or clean the filter. Only use filters supplied by your distributor. And do not operate without a filter.

- 2. Change the filter every 30 days or when the filter turn gray.
- 3. Changing procedure: a. Remove the filter cover.

Air Opening

b. Replace the used filter with a new one. C. Replace the filter cover.

6. REPLACEMENT PARTS & OPTIONAL ACCESSORIES

Replacement parts are available through your local retailer or they can be ordered by mail. Optional accessories include adult mask and baby pacifier inhaler. The pacifier inhaler allows baby to receive aerosol treatment easily through normal breath.

# 7. SPECIFICATIONS

Input Voltage	AC220V, 50Hz
Power	45 watt
Current	0.5A
Medication Capacity	8ml.
Particle Size	0.5 to 10μm
MMAD	4µm
Sound Level	Less than 60dB at 1 meter
Average Nebulizer Rate	Min. 0.2ml/min.
Compressor Pressure Range	32-45Psi
Operating Pressure Range	16 to 22 Psi
Liter Flow Range	6-9Lpm
Weight	1.6Kg
Storage Conditions	Keep it in cool & dry place
Shelf-life of Nebulizer Kit	5 years
Warranty	1 Year

Protection against electric shock


□ - Class II Equipment

Degree of safety in the presence of flammable anesthetics or oxygen

-No AP I APG (Not suitable for use in the presence of flammable anesthetics  $^{\mbox{or}}$ 


IPXO - Protection against harmful in gress of water is Ordinary. Mode of operation-Continuous.


Compressor nebulizer the particle size distribution curve as well: Result of cascade impactor \*\*measurements for particle size with Compressor



Particle Size Dp (um) MMAD = Mass Median Aerodynamic Diameter

Once the product life has ended or its components do not work anymore their disposal should be carried out according to the current regulations





workmanship and operation, under normal use, for a period of one year from the date of purchase. If any part become defective during the warranty period Passim Lifesciences Ltd. will repair or replace (if not repairable) the same, free of cost. This

shall not apply to any parts that are considered as expandible or deteriorable in the course of normal use. Passim Lifesciences Ltd. shall be relieved of any liability and warranty shall cease to apply if:

- This is not used in accordance with the instructions in the operational manual.
- It is used with any equipment not complying with the specification of this unit.
- It is not regularly maintained. • The unit is disassembled, repaired or operated by person not authorized by
- Passim Lifesciences Ltd.
- · Damage caused due to negligency.
- The unit is operated in corrosive materials or in the harmful atmosphere. • The warranty card is not filled completely and produced at the time of warranty

# Manufactured by:

Passim Medichem Agencies At: A-39, Naraina Industrial Area, Phase-II. New Delhi - 110028 (INDIA) Loan Lic. No.: MFG/MD/2022/000405 Please Contact: 1800 309 3009 (Timing: 9am -7pm, Mon. - Sat.)

Email ID: customercare@drodin.ir

Website: www.drodin.in

